

20 Progress Drive Morrisville, PA 19067 215-736-1126 215-736-1128 [fax]

Product Data Sheet

Product PP2-0900RX

ColorRx®

Biocompatible, Polypropylene random copolymer with radiation resistance for healthcare applications.

PHYSICAL	Test Method	Typical Values, Units	7/3/2012	
Specific Gravity	ASTM D792	.902 g/cm 3		
Melt Flow Rate 190C / 2.16 kg	ASTM D1238	9.0 g/10 min		
Mold Shrinkage Linear Flow (0.125)		in/in		
Water Absorption @ 24 hrs		%		
IMPACT	Test Method	Typical Values, Units		
Izod Impact Strength	ASTM D256	1.1 ft-lb/in		
Notched (73 F) (-22 F)		ft-lb/in		
MECHANICAL	Test Method	Typical Values, Units		
Tensile Strength @ Yield**	ASTM D638	3960 psi		
Tensile Strength @ Break**		psi		
Elongation @ Yield*		%		
Elongation @ Break*	ASTM D638	>700 %		
Flexural Strength***		psi		
Flexural Modulus***		psi		
HARDNESS	Test Method	Typical Values, Units		
Hardness R Scale	ASTM D785	90		
THERMAL	Test Method	Typical Values, Units		
DTUL @ 66 psi	ASTM D648	194 °F		
Unannealed				
IGNITION CHARACTERISTICS	Test Method	Typical Values, Units		

UL File Number Flame Rating - UL

The values shown are typical values that have been obtained using test bars molded from laboratory samples and are not intended for specification purposes. These values are for natural colors only. Addition of pigments may alter some values. Inasmuch as LTL Color Compounders has no control over the use to which others may put the material, it does not guarantee that the same results as those described here in will be obtained. Each user of the material should make his own test to detail the material's suitable user. particular use. Statements concerning possible or suggested uses of the materials described herein are not to be construed as constituting a license under any LTL Color Compounders patent covering such use or as recommendations for use of such materials in the infringement of any patent. These are developmental products with estimated physical property profiles. Actual values will need to be determined upon production of material.

^{* %} elongation values are calculated from the elongation of the entire bar at 2.0 in/min

^{***} Tensile strength values are calculated at 2.0 in/min
*** Flexural data is calculated at 2.0 in/min